

AECOM Asia Company Limited

TSP - Total Suspended Particulates Sampler Field Calibration Report

 Station Kwun Tong Government Secondary School (ID1A)

 Operator: Choi Wing Ho

 Date: 24-Dec-13

 Next Due Date: 24-Feb-14

 Pump No.: 763

 Verified Against: O.T.S. -- 0988

 Equipment No.: A-001-64T

 Expiration Date: 20-May-14

| Ambient Condition | | | | |
|-------------------|-----|--------|--------------|------------|
| Temperature, Ta | 289 | Kelvin | Pressure, Pa | 756.9 mmHg |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|-----------|---|---------|---------------|---------|
| Equipment No.: | 988 | Slope, mc | 1.94727 | Intercept, bc | 0.02332 |
| Last Calibration Date: | 20-May-13 | $mc \times Qstd + bc = [H \times (Pa/760) \times (298/Ta)]^{1/2}$ | | | |
| Next Calibration Date: | 20-May-14 | | | | |

| Calibration of TSP Sampler | | | | | |
|----------------------------|-------------------|--|---|-----------------|---|
| Calibration Point | H in. of water | [H x (Pa/760) x (298/Ta)] ^{1/2} | Qstd (m ³ /min) X - axis | W in. of oil | [ΔW x (Pa/760) x (298/Ta)] ^{1/2} Y-axis |
| 1 | 9.0 | 3.04 | 1.55 | 5.8 | 2.44 |
| 2 | 8.1 | 2.88 | 1.47 | 5.0 | 2.27 |
| 3 | 6.4 | 2.56 | 1.30 | 3.5 | 1.90 |
| 4 | 4.3 | 2.10 | 1.07 | 2.2 | 1.50 |
| 5 | 2.8 | 1.70 | 0.86 | 1.3 | 1.16 |

By Linear Regression of Y on X

 Slope, mw = 1.8562

 Intercept, bw = -0.4662

 Correlation Coefficient* = 0.9980

Set Point Calculation

 From the TSP Field Calibration Curve, take Qstd = 1.21 m³/min (43 CFM)

From the Regression Equation, the "Y" value according to

$$m \times Qstd + b = [W \times (Pa/760) \times (298/Ta)]^{1/2}$$

 Therefore, Set Point W = $(m \times Qstd + b)^2 \times (760 / Pa) \times (Ta / 298) =$ 3.08

*If Correlation Coefficient < 0.990, check and recalibrate again.

 Remarks: _____

 QC Reviewer: YI Leung

 Signature: [Signature]

 Date: 27-12-13

AECOM Asia Company Limited

TSP - Total Suspended Particulates Sampler Field Calibration Report

 Station Sau Mau Ping Catholic Primary School (ID5)

 Operator: Shum Kam Yuen

 Date: 18-Dec-13

 Next Due Date: 18-Feb-14

 Pump No.: 10088

 Verified Against: O.T.S. -- 0988

 Equipment No.: A-001-13T

 Expiration Date: 20-May-14

| Ambient Condition | | | | |
|-------------------|-----|--------|--------------|------------|
| Temperature, Ta | 289 | Kelvin | Pressure, Pa | 756.9 mmHg |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|-----------|---|---------|---------------|---------|
| Equipment No.: | 988 | Slope, mc | 1.94727 | Intercept, bc | 0.02332 |
| Last Calibration Date: | 20-May-13 | $mc \times Qstd + bc = [H \times (Pa/760) \times (298/Ta)]^{1/2}$ | | | |
| Next Calibration Date: | 20-May-14 | | | | |

| Calibration of TSP Sampler | | | | | |
|----------------------------|-------------------|--|---|-----------------|---|
| Calibration Point | H in. of water | [H x (Pa/760) x (298/Ta)] ^{1/2} | Qstd (m ³ /min) X - axis | W in. of oil | [ΔW x (Pa/760) x (298/Ta)] ^{1/2} Y-axis |
| 1 | 8.8 | 3.01 | 1.53 | 6.2 | 2.52 |
| 2 | 7.6 | 2.79 | 1.42 | 5.0 | 2.27 |
| 3 | 5.3 | 2.33 | 1.18 | 3.5 | 1.90 |
| 4 | 4 | 2.03 | 1.03 | 2.6 | 1.63 |
| 5 | 2.8 | 1.70 | 0.86 | 1.6 | 1.28 |

By Linear Regression of Y on X

 Slope, mw = 1.7950

 Intercept, bw = -0.2412

 Correlation Coefficient* = 0.9984
Set Point Calculation

 From the TSP Field Calibration Curve, take Qstd = 1.21 m³/min (43 CFM)

From the Regression Equation, the "Y" value according to

$$m \times Qstd + b = [W \times (Pa/760) \times (298/Ta)]^{1/2}$$

 Therefore, Set Point W = (m x Qstd + b)² x (760 / Pa) x (Ta / 298) = 3.63

*If Correlation Coefficient < 0.990, check and recalibrate again.

 Remarks: _____

 QC Reviewer: YI Lang

 Signature: 

 Date: 27-12-13

AECOM Asia Company Limited

TSP - Total Suspended Particulates Sampler Field Calibration Report

 Station Kwun Tong Government Secondary School (ID1A)

 Operator: Leung Yiu Ting

 Date: 19-Feb-14

 Next Due Date: 19-May-14

 Pump No.: 763

 Verified Against: O.T.S. -- 0988

 Equipment No.: A-001-64T

 Expiration Date: 20-May-14

| Ambient Condition | | | | | |
|-------------------|-----|--------|--------------|-------|------|
| Temperature, Ta | 283 | Kelvin | Pressure, Pa | 766.2 | mmHg |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|-----------|---|---------|---------------|---------|
| Equipment No.: | 988 | Slope, mc | 1.94727 | Intercept, bc | 0.02332 |
| Last Calibration Date: | 20-May-13 | $mc \times Qstd + bc = [H \times (Pa/760) \times (298/Ta)]^{1/2}$ | | | |
| Next Calibration Date: | 20-May-14 | | | | |

| Calibration of TSP Sampler | | | | | |
|----------------------------|-------------------|--|---|-----------------|---|
| Calibration Point | H in. of water | [H x (Pa/760) x (298/Ta)] ^{1/2} | Qstd (m ³ /min) X - axis | W in. of oil | [ΔW x (Pa/760) x (298/Ta)] ^{1/2} Y-axis |
| 1 | 8.9 | 3.07 | 1.56 | 5.9 | 2.50 |
| 2 | 8.0 | 2.91 | 1.48 | 5.1 | 2.33 |
| 3 | 6.2 | 2.57 | 1.31 | 3.5 | 1.93 |
| 4 | 4.4 | 2.16 | 1.10 | 2.3 | 1.56 |
| 5 | 2.9 | 1.75 | 0.89 | 1.4 | 1.22 |

By Linear Regression of Y on X

 Slope, mw = 1.9196

Intercept, bw =

-0.5260

 Correlation Coefficient* = 0.9970

Set Point Calculation

 From the TSP Field Calibration Curve, take Qstd = 1.21 m³/min (43 CFM)

From the Regression Equation, the "Y" value according to

$$m \times Qstd + b = [W \times (Pa/760) \times (298/Ta)]^{1/2}$$

 Therefore, Set Point W = (m x Qstd + b)² x (760 / Pa) x (Ta / 298) = 3.04

*If Correlation Coefficient < 0.990, check and recalibrate again.

 Remarks: _____

 QC Reviewer: HY Sun

 Signature: Yy

 Date: 21-2-14

TSP - Total Suspended Particulates Sampler Field Calibration Report

Station Sau Nga House (ID3)
Date: 19-Feb-14
Pump No.: 1272
Equipment No.: A-001-31T

Operator: Leung Yiu Ting
Next Due Date: 19-May-14
Verified Against: O.T.S. -- 0988
Expiration Date: 20-May-14

| Ambient Condition | | | | | |
|-------------------|-----|--------|--------------|-------|------|
| Temperature, Ta | 283 | Kelvin | Pressure, Pa | 766.2 | mmHg |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|-----------|---|---------|---------------|---------|
| Equipment No.: | 988 | Slope, mc | 1.94727 | Intercept, bc | 0.02332 |
| Last Calibration Date: | 20-May-13 | $mc \times Qstd + bc = [H \times (Pa/760) \times (298/Ta)]^{1/2}$ | | | |
| Next Calibration Date: | 20-May-14 | | | | |

| Calibration of TSP Sampler | | | | | |
|----------------------------|----------------|---|--|--------------|--|
| Calibration Point | H in. of water | $[H \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m ³ /min) X - axis | W in. of oil | $[\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$ Y-axis |
| 1 | 9.0 | 3.09 | 1.57 | 6.6 | 2.65 |
| 2 | 7.2 | 2.76 | 1.41 | 4.5 | 2.19 |
| 3 | 5.8 | 2.48 | 1.26 | 3.5 | 1.93 |
| 4 | 4.4 | 2.16 | 1.10 | 2.5 | 1.63 |
| 5 | 3.5 | 1.93 | 0.98 | 1.7 | 1.34 |

By Linear Regression of Y on X
Slope, mw = 2.1348 Intercept, bw = -0.7504
Correlation Coefficient* = 0.9959

Set Point Calculation

From the TSP Field Calibration Curve, take Qstd = 1.21 m³/min (43 CFM)
From the Regression Equation, the "Y" value according to

$$m \times Qstd + b = [W \times (Pa/760) \times (298/Ta)]^{1/2}$$

Therefore, Set Point W = $(m \times Qstd + b)^2 \times (760 / Pa) \times (Ta / 298) =$ 3.16

*If Correlation Coefficient < 0.990, check and recalibrate again.

Remarks: _____

QC Reviewer: HY Siu

Signature: ly

Date: 21-2-14



TISCH ENVIRONMENTAL, INC.
 145 SOUTH MIAMI AVE.
 VILLAGE OF CLEVELAND, OH 45002
 513.467.9000
 877.263.7610 TOLL FREE
 513.467.9009 FAX
 WWW.TISCH-ENV.COM

AIR POLLUTION MONITORING EQUIPMENT

ORIFICE TRANSFER STANDARD CERTIFICATION WORKSHEET TE-5025A

Date - May 20, 2013 Rootsometer S/N 0438320 Ta (K) - 297
 Operator Tisch Orifice I.D. - 0988 Pa (mm) - 751.84

| PLATE OR Run # | VOLUME START (m3) | VOLUME STOP (m3) | DIFF VOLUME (m3) | DIFF TIME (min) | METER DIFF Hg (mm) | ORFICE DIFF H2O (in.) |
|----------------|-------------------|------------------|------------------|-----------------|--------------------|-----------------------|
| 1 | NA | NA | 1.00 | 1.3900 | 3.2 | 2.00 |
| 2 | NA | NA | 1.00 | 0.9720 | 6.4 | 4.00 |
| 3 | NA | NA | 1.00 | 0.8670 | 7.9 | 5.00 |
| 4 | NA | NA | 1.00 | 0.8270 | 8.7 | 5.50 |
| 5 | NA | NA | 1.00 | 0.6800 | 12.6 | 8.00 |

DATA TABULATION

| Vstd | (x axis) Qstd | (y axis) | Va | (x axis) Qa | (y axis) |
|------------------------------------|---------------|----------|---------------------------|-------------|----------|
| 0.9884 | 0.7110 | 1.4090 | 0.9957 | 0.7163 | 0.8889 |
| 0.9842 | 1.0125 | 1.9926 | 0.9915 | 1.0201 | 1.2570 |
| 0.9821 | 1.1327 | 2.2278 | 0.9894 | 1.1412 | 1.4054 |
| 0.9811 | 1.1863 | 2.3365 | 0.9884 | 1.1952 | 1.4740 |
| 0.9759 | 1.4352 | 2.8179 | 0.9832 | 1.4459 | 1.7777 |
| Qstd slope (m) = 1.94727 | | | Qa slope (m) = 1.21935 | | |
| intercept (b) = 0.02332 | | | intercept (b) = 0.01471 | | |
| coefficient (r) = 0.99998 | | | coefficient (r) = 0.99998 | | |
| y axis = SQRT[H2O(Pa/760)(298/Ta)] | | | y axis = SQRT[H2O(Ta/Pa)] | | |

CALCULATIONS

Vstd = Diff. Vol [(Pa-Diff. Hg)/760] (298/Ta)
 Qstd = Vstd/Time

Va = Diff Vol [(Pa-Diff Hg)/Pa]
 Qa = Va/Time

For subsequent flow rate calculations:

Qstd = 1/m{ [SQRT(H2O(Pa/760)(298/Ta))] - b}
 Qa = 1/m{ [SQRT H2O(Ta/Pa)] - b}

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3
 Equipment No.: A.005.07a
 Sensitivity Adjustment Scale Setting: 557 CPM

Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No: Control: 140AB219899803
 Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 18 May 2013

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 557 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 557 CPM

| Hour | Date (dd-mm-yy) | Time | Ambient Condition | | Concentration ¹ (mg/m ³) Y-axis | Total Count ² | Count/ Minute ³ X-axis |
|------|--------------------|---------------|-------------------|-------------|--|--------------------------|---|
| | | | Temp (°C) | R.H. (%) | | | |
| 1 | 18-05-13 | 12:30 - 13:30 | 28.1 | 78 | 0.04714 | 1887 | 31.45 |
| 2 | 18-05-13 | 13:30 - 14:30 | 28.1 | 78 | 0.04932 | 1970 | 32.83 |
| 3 | 18-05-13 | 14:30 - 15:30 | 28.2 | 77 | 0.05156 | 2056 | 34.27 |
| 4 | 18-05-13 | 15:30 - 16:30 | 28.1 | 78 | 0.05083 | 2026 | 33.77 |

- Note:
1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015
 Correlation coefficient: 0.9978

Validity of Calibration Record: 17 May 2014

Remarks:

QC Reviewer: YW Fung Signature:  Date: 20 May 2013

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3
 Equipment No.: A.005.08a
 Sensitivity Adjustment Scale Setting: 702 CPM

 Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No: Control: 140AB219899803
 Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 18 May 2013

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 702 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 702 CPM

| Hour | Date (dd-mm-yy) | Time | Ambient Condition | | Concentration ¹ (mg/m ³) Y-axis | Total Count ² | Count/ Minute ³ X-axis |
|------|--------------------|---------------|-------------------|-------------|--|-----------------------------|---|
| | | | Temp (°C) | R.H. (%) | | | |
| 1 | 18-05-13 | 12:30 - 13:30 | 28.1 | 78 | 0.04714 | 1764 | 29.40 |
| 2 | 18-05-13 | 13:30 - 14:30 | 28.1 | 78 | 0.04932 | 1846 | 30.77 |
| 3 | 18-05-13 | 14:30 - 15:30 | 28.2 | 77 | 0.05156 | 1935 | 32.25 |
| 4 | 18-05-13 | 15:30 - 16:30 | 28.1 | 78 | 0.05083 | 1899 | 31.65 |

- Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0016
 Correlation coefficient: 0.9976

Validity of Calibration Record: 17 May 2014

Remarks:

QC Reviewer: YW Fung Signature:  Date: 20 May 2013

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3
 Equipment No.: A.005.09a
 Sensitivity Adjustment Scale Setting: 797 CPM

Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No: Control: 140AB219899803
 Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 18 May 2013

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 797 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 797 CPM

| Hour | Date (dd-mm-yy) | Time | Ambient Condition | | Concentration ¹ (mg/m ³) Y-axis | Total Count ² | Count/ Minute ³ X-axis |
|------|--------------------|---------------|-------------------|-------------|--|--------------------------|---|
| | | | Temp (°C) | R.H. (%) | | | |
| 1 | 18-05-13 | 12:30 - 13:30 | 28.1 | 78 | 0.04714 | 1885 | 31.42 |
| 2 | 18-05-13 | 13:30 - 14:30 | 28.1 | 78 | 0.04932 | 1965 | 32.75 |
| 3 | 18-05-13 | 14:30 - 15:30 | 28.2 | 77 | 0.05156 | 2059 | 34.32 |
| 4 | 18-05-13 | 15:30 - 16:30 | 28.1 | 78 | 0.05083 | 2024 | 33.73 |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015
 Correlation coefficient: 0.9973

Validity of Calibration Record: 17 May 2014

Remarks:

QC Reviewer: YW Fung Signature:  Date: 20 May 2013

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3
 Equipment No.: A.005.10a
 Sensitivity Adjustment Scale Setting: 753 CPM

Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No.: Control: 140AB219899803
Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 18 May 2013

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 753 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 753 CPM

| Hour | Date (dd-mm-yy) | Time | Ambient Condition | | Concentration ¹ (mg/m ³) Y-axis | Total Count ² | Count/ Minute ³ X-axis |
|------|--------------------|---------------|-------------------|-------------|--|--------------------------|---|
| | | | Temp (°C) | R.H. (%) | | | |
| 1 | 18-05-13 | 12:30 - 13:30 | 28.1 | 78 | 0.04714 | 1886 | 31.43 |
| 2 | 18-05-13 | 13:30 - 14:30 | 28.1 | 78 | 0.04932 | 1968 | 32.80 |
| 3 | 18-05-13 | 14:30 - 15:30 | 28.2 | 77 | 0.05156 | 2061 | 34.35 |
| 4 | 18-05-13 | 15:30 - 16:30 | 28.1 | 78 | 0.05083 | 2026 | 33.77 |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015
 Correlation coefficient: 0.9983

Validity of Calibration Record: 17 May 2014

Remarks:

QC Reviewer: YW Fung Signature:  Date: 20 May 2013

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3
 Equipment No.: A.005.11a
 Sensitivity Adjustment Scale Setting: 799 CPM

Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No: Control: 140AB219899803
 Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 18 May 2013

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 799 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 799 CPM

| Hour | Date (dd-mm-yy) | Time | Ambient Condition | | Concentration ¹ (mg/m ³) Y-axis | Total Count ² | Count/ Minute ³ X-axis |
|------|--------------------|---------------|-------------------|-------------|--|--------------------------|---|
| | | | Temp (°C) | R.H. (%) | | | |
| 1 | 18-05-13 | 12:15 - 13:15 | 28.1 | 78 | 0.04685 | 1871 | 31.18 |
| 2 | 18-05-13 | 13:15 - 14:15 | 28.1 | 78 | 0.04941 | 1979 | 32.98 |
| 3 | 18-05-13 | 14:15 - 15:15 | 28.2 | 77 | 0.05127 | 2055 | 34.25 |
| 4 | 18-05-13 | 15:15 - 16:15 | 28.1 | 78 | 0.05060 | 2021 | 33.68 |


Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015
 Correlation coefficient: 0.9976

Validity of Calibration Record: 17 May 2014

Remarks:

QC Reviewer: YW Fung Signature:  Date: 20 May 2013

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3B
 Equipment No.: A.005.13a
 Sensitivity Adjustment Scale Setting: 643 CPM

Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No: Control: 140AB219899803
 Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 18 May 2013

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 643 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 643 CPM

| Hour | Date (dd-mm-yy) | Time | Ambient Condition | | Concentration ¹ (mg/m ³) Y-axis | Total Count ² | Count/ Minute ³ X-axis |
|------|--------------------|---------------|-------------------|-------------|--|--------------------------|---|
| | | | Temp (°C) | R.H. (%) | | | |
| 1 | 18-05-13 | 12:15 - 13:15 | 28.1 | 78 | 0.04685 | 1867 | 31.12 |
| 2 | 18-05-13 | 13:15 - 14:15 | 28.1 | 78 | 0.04941 | 1975 | 32.92 |
| 3 | 18-05-13 | 14:15 - 15:15 | 28.2 | 77 | 0.05127 | 2048 | 34.13 |
| 4 | 18-05-13 | 15:15 - 16:15 | 28.1 | 78 | 0.05060 | 2017 | 33.62 |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015
 Correlation coefficient: 0.9986

Validity of Calibration Record: 17 May 2014

Remarks:

QC Reviewer: YW Fung Signature:  Date: 20 May 2013

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3B
 Equipment No.: A.005.14a
 Sensitivity Adjustment Scale Setting: 786 CPM

 Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No: Control: 140AB219899803
 Sensor: 1200C143659803 K_o: 12500
 Last Calibration Date*: 18 May 2013

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 786 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 786 CPM

| Hour | Date (dd-mm-yy) | Time | Ambient Condition | | Concentration ¹ (mg/m ³) Y-axis | Total Count ² | Count/ Minute ³ X-axis |
|------|--------------------|---------------|-------------------|-------------|--|--------------------------|---|
| | | | Temp (°C) | R.H. (%) | | | |
| 1 | 18-05-13 | 12:15 - 13:15 | 28.1 | 78 | 0.04685 | 2005 | 33.42 |
| 2 | 18-05-13 | 13:15 - 14:15 | 28.1 | 78 | 0.04941 | 2121 | 35.35 |
| 3 | 18-05-13 | 14:15 - 15:15 | 28.2 | 77 | 0.05127 | 2194 | 36.57 |
| 4 | 18-05-13 | 15:15 - 16:15 | 28.1 | 78 | 0.05060 | 2167 | 36.12 |

- Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0014
 Correlation coefficient: 0.9987

Validity of Calibration Record: 17 May 2014

Remarks:

QC Reviewer: YW Fung Signature:  Date: 20 May 2013

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3B
 Equipment No.: A.005.16a
 Sensitivity Adjustment Scale Setting: 521 CPM
 Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No.: Control: 140AB219899803
 Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 18 May 2013

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 521 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 521 CPM

| Hour | Date (dd-mm-yy) | Time | Ambient Condition | | Concentration ¹ (mg/m ³) Y-axis | Total Count ² | Count/ Minute ³ X-axis |
|------|--------------------|---------------|-------------------|-------------|---|--------------------------|--|
| | | | Temp (°C) | R.H. (%) | | | |
| 1 | 27-07-13 | 11:00 - 12:00 | 27.3 | 75 | 0.04734 | 1893 | 31.55 |
| 2 | 27-07-13 | 12:00 - 13:00 | 27.3 | 75 | 0.04789 | 1915 | 31.92 |
| 3 | 27-07-13 | 13:00 - 14:00 | 27.4 | 74 | 0.04953 | 1976 | 32.93 |
| 4 | 27-07-13 | 14:00 - 15:00 | 27.4 | 75 | 0.04867 | 1949 | 32.48 |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015
 Correlation coefficient: 0.9934

Validity of Calibration Record: 26 July 2014

Remarks:

QC Reviewer: YW Fung Signature:  Date: 29 July 2013



CERTIFICATE OF CALIBRATION

Certificate No.: 13CA1107 01-01 Page 1 of 2

Item tested

| | | | |
|-----------------------|----------------------------|---|----------------|
| Description: | Sound Level Meter (Type 1) | , | Microphone |
| Manufacturer: | Rion Co., Ltd. | , | Rion Co., Ltd. |
| Type/Model No.: | NL-31 | , | UC-53A |
| Serial/Equipment No.: | 00320528 / N.007.03A | , | 90565 |
| Adaptors used: | - | , | - |

Item submitted by

Customer Name: AECOM ASIA CO., LTD.
Address of Customer: -
Request No.: -
Date of receipt: 07-Nov-2013

Date of test: 08-Nov-2013

Reference equipment used in the calibration

| Description: | Model: | Serial No. | Expiry Date: | Traceable to: |
|---------------------------------|----------|------------|--------------|---------------|
| Multi function sound calibrator | B&K 4226 | 2288444 | 22-Jun-2014 | CIGISMEC |
| Signal generator | DS 360 | 33873 | 15-Apr-2014 | CEPREI |
| Signal generator | DS 360 | 61227 | 15-Apr-2014 | CEPREI |

Ambient conditions

Temperature: 22 ± 1 °C
Relative humidity: 60 ± 10 %
Air pressure: 1000 ± 10 hPa

Test specifications

- 1, The Sound Level Meter has been calibrated in accordance with the requirements as specified in BS 7580: Part 1: 1997 and the lab calibration procedure SMTP004-CA-152.
- 2, The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of ±20%.
- 3, The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure responsiveness of the Sound Level Meter.

Test results

This is to certify that the Sound Level Meter conforms to BS 7580: Part 1: 1997 for the conditions under which the test was performed.

Details of the performed measurements are presented on page 2 of this certificate.

Actual Measurement data are documented on worksheets.

Approved Signatory:

Huang Jian Min/Feng Jun Qi

Date: 11-Nov-2013

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.



CERTIFICATE OF CALIBRATION

Certificate No.: 13CA0305 01-01 Page 1 of 2

Item tested

| | | | |
|-----------------------|----------------------------|---|------------|
| Description: | Sound Level Meter (Type 1) | , | Microphone |
| Manufacturer: | B & K | , | B & K |
| Type/Model No.: | 2250-L | , | 4950 |
| Serial/Equipment No.: | 2681366 (N-011.01) | , | 2665582 |
| Adaptors used: | - | , | - |

Item submitted by

Customer Name: AECOM ASIA CO LIMITED
Address of Customer: -
Request No.: -
Date of receipt: 05-Mar-2013

Date of test: 05-Mar-2013

Reference equipment used in the calibration

| Description: | Model: | Serial No. | Expiry Date: | Traceable to: |
|---------------------------------|----------|------------|--------------|---------------|
| Multi function sound calibrator | B&K 4226 | 2288444 | 23-May-2013 | CIGISMEC |
| Signal generator | DS 360 | 33873 | 29-May-2013 | CEPREI |
| Signal generator | DS 360 | 61227 | 29-May-2013 | CEPREI |

Ambient conditions

Temperature: 21 ± 1 °C
Relative humidity: 60 ± 10 %
Air pressure: 1000 ± 10 hPa

Test specifications

- 1, The Sound Level Meter has been calibrated in accordance with the requirements as specified in BS 7580: Part 1: 1997 and the lab calibration procedure SMTP004-CA-152.
- 2, The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of $\pm 20\%$.
- 3, The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure responsiveness of the Sound Level Meter.

Test results

This is to certify that the Sound Level Meter conforms to BS 7580: Part 1: 1997 for the conditions under which the test was performed.

Details of the performed measurements are presented on page 2 of this certificate.

Actual Measurement data are documented on worksheets.

Approved Signatory:

Huang Jian Min/Feng Jun Qi

Date: 05-Mar-2013

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.



CERTIFICATE OF CALIBRATION

Certificate No.: 13CA1107 01-02

Page: 1 of 2

Item tested

Description: Acoustical Calibrator (Class 1)
Manufacturer: Rion Co., Ltd.
Type/Model No.: NC-73
Serial/Equipment No.: 10307223 / N.004.08
Adaptors used: -

Item submitted by

Customer: AECOM ASIA CO., LTD.
Address of Customer: -
Request No.: -
Date of receipt: 07-Nov-2013

Date of test: 08-Nov-2013

Reference equipment used in the calibration

| Description: | Model: | Serial No. | Expiry Date: | Traceable to: |
|-------------------------|----------|------------|--------------|---------------|
| Lab standard microphone | B&K 4180 | 2341427 | 17-Apr-2014 | SCL |
| Preamplifier | B&K 2673 | 2239857 | 16-Apr-2014 | CEPREI |
| Measuring amplifier | B&K 2610 | 2346941 | 24-Apr-2014 | CEPREI |
| Signal generator | DS 360 | 61227 | 15-Apr-2014 | CEPREI |
| Digital multi-meter | 34401A | US36087050 | 10-Dec-2013 | CEPREI |
| Audio analyzer | 8903B | GB41300350 | 15-Apr-2014 | CEPREI |
| Universal counter | 53132A | MY40003662 | 15-Apr-2014 | CEPREI |

Ambient conditions

Temperature: 22 ± 1 °C
Relative humidity: 60 ± 10 %
Air pressure: 1000 ± 10 hPa

Test specifications

- 1, The Sound Calibrator has been calibrated in accordance with the requirements as specified in IEC 60942 1997 Annex B and the lab calibration procedure SMTP004-CA-156.
- 2, The calibrator was tested with its axis vertical facing downwards at the specific frequency using insert voltage technique.
- 3, The results are rounded to the nearest 0.01 dB and 0.1 Hz and have not been corrected for variations from a reference pressure of 1013.25 hectoPascals as the maker's information indicates that the instrument is insensitive to pressure changes.

Test results

This is to certify that the sound calibrator conforms to the requirements of annex B of IEC 60942: 1997 for the conditions under which the test was performed. This does not imply that the sound calibrator meets IEC 60942 under any other conditions.

Details of the performed measurements are presented on **page 2** of this certificate.

Approved Signatory:


Huang Jian Min/Feng Jun Qi

Date: 11-Nov-2013

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.